

How To Upgrade RadSight Terminal Software

Document Part Number: CDM15214

Issue: 1.0

Date Issued: 17th June 2025

© 2025 Charthouse Data Management Ltd.

All rights reserved.

All material in this document is protected by copyright. No part of this content may be reproduced in any form by any means without prior written authorization of Charthouse Data Management Ltd. The information described in this document may be protected by one or more UK patents, foreign patents, or pending applications.

Trademarks

Product names mentioned herein are the trademarks of their respective owners.

Change Log

VERSION	DATE	AUTHOR	COMMENTS
1.0	17 th June 2025	David Kuziara	First Draft

This document was produced by Charthouse Data Management Ltd. It may contain privileged and confidential information intended solely for the use of Charthouse Data Management Ltd. and the customers, suppliers or partners of Charthouse Data Management Ltd. specifically associated with the matters addressed herein. If you are not the intended recipient, you should immediately stop reading this document and destroy it. Any unauthorized reading, distribution, copying, or other use of this document is strictly prohibited. If you received this document in error, please notify Charthouse Data Management Ltd. immediately by calling +44 1202 041444

CONTENTS

1 Overview.....4

2 Steps to upgrade RadSight using the script.....4

1 OVERVIEW

Versions of RadSight lower than version 2.60 installed on Terminals will now need to be upgraded using an upgrade script. This document takes you through the steps to do this.

2 STEPS TO UPGRADE RADSIGHT USING THE SCRIPT

- You will need a keyboard and mouse (ideally using a USB hub)
- Log in as the Administrator account
 - Credentials can be found on the card that came with RadSight
 - To get to the user account choice:
 - ctrl+alt+del
 - or press the Windows key 5 times
- Click the Windows key
- Type **cmd** and run it as administrator
- Type **D:** and press return
- Type **CD Win10** and press return
- Type **powershell -ExecutionPolicy Bypass -File ".\Terminal-Upgrade-UWP-to-WPF.ps1"** and press return
- If prompted, click **Yes** (to acknowledge running as administrator)
 - The screen may jump back to desktop
 - if so, hold down the Windows key and the Tab key to select the PowerShell screen that asks for the location of the RadSight installer
- When prompted for the location of the RadSight installer, it will default to **D:\Win10** so just hit return
 - The script will take a few minutes to process
 - The screen may go back to the desktop – but it is still upgrading
 - Wait a few minutes
 - The terminal will then reboot

Authored By: David Kuziara

Date: 12th June 2025

Approved By:

Date:

This document was produced by Charthouse Data Management Ltd. It may contain privileged and confidential information intended solely for the use of Charthouse Data Management Ltd. and the customers, suppliers or partners of Charthouse Data Management Ltd. specifically associated with the matters addressed herein. If you are not the intended recipient, you should immediately stop reading this document and destroy it. Any unauthorized reading, distribution, copying, or other use of this document is strictly prohibited. If you received this document in error, please notify Charthouse Data Management Ltd. immediately by calling 01202 031444.